



**CERTIFICATE OF ANALYSIS**

BioProfile Control Level 9  
Lot Number 25246010

Catalog Number 43473  
Exp Date 2026-12

	Spec	Pass/Fail
Particles/mL	55.00 x 10 <sup>5</sup> - 65.00 x 10 <sup>5</sup>	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "C. S. D. E. A.", is written over a horizontal line.

Approval

QC Inspector  
Title

10/29/2025  
Date