



**CERTIFICATE OF ANALYSIS**

STP pHox Plus/L/C/M Ca Sensor  
Lot Number 25281036

Part Number 27813  
Exp Date 2027-01-08  
Date of Manufacture 2025-10-08  
UDI#(01)00385480278133(11)251008(17)270108(10)25281036

	<u>Spec*</u>	<u>Pass/Fail</u>
Ca Slope	≥ 9.30	Pass
Std A mV	0 ± 60.00	Pass
pHOx Ctl L1	1.37 - 1.57	Pass
pHOx Ctl L3	0.50 - 0.62	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

\*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

10/31/2025

Date