



CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pH Sensor
Lot Number 25282032

Ref Number 21522
Exp Date 2027-10-09
Date of Manufacture 2025-10-09
UDI# (01)00385480215220(11)251009(17)271009(10)25282032

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	10.0 - 10.8	Pass
Std A mV	0.00 +/- 110.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector

Title

10/23/2025

Date