



**CERTIFICATE OF ANALYSIS**

STAT PROFILE pHox Ultra/CCX Potassium Sensor  
Lot Number 25281039

Ref Number 48582

Exp Date 2027-01-08

Date of Manufacture 2025-10-08

UDI# (01)00385480485821(11)251008(17)270108(10)25281039

|          | <u>Spec</u>   | <u>Pass/Fail</u> |
|----------|---------------|------------------|
| Slope    | 9.00 to 11.00 | Pass             |
| Std C mV | 0 ± 50.00     | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Z. Veramwala*  
Approval

QC Inspector  
Title

11/4/2025  
Date