



CERTIFICATE OF ANALYSIS

BioProfile Cell Density Calibrator
Lot Number 25272009

Ref Number 43034
Exp Date 2026-12

	<u>Spec</u>	<u>Pass/Fail</u>
Particles/mL	23.0 x 10 ⁵ - 27.0 x 10 ⁵	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector
Title

11/5/2025
Date