



**CERTIFICATE OF ANALYSIS**

BioPro Flex Deproteinizing Solution  
Lot Number 25290021

Ref Number 46850  
Exp Date 2026-10

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl %	0.70 - 0.90	Pass
pH	9.00 - 11.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "John", is written over a horizontal line.

Approval

QC Inspector

Title

11/5/2025

Date