



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 CDV Reagent Pack
Lot Number 25303031

Ref Number 55297
Exp Date 2026-09-22
Date of Manufacture 2025-09-22

	CDV Bottle Pack L/N:	25303030	Exp Date:	2026-09-22		
Soln SF L/N	25265022		Soln DP L/N	25265023		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>	
pH	6.900 - 7.100	Pass	Depro Soln	N/A	Present	
Cal C L/N	25265063		Soln T L/N	25265024		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>	
3% Detergent	N/A	Present	0.4% Trypan	N/A	Present	

CDV Reagent Cartridge L/N:	25303028	Exp Date:	2026-12-25
Soln PF L/N	25268021		
	<u>Spec</u>	<u>Pass/Fail</u>	
pH	6.900 - 7.100	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector
Title

11/10/2025
Date