



CERTIFICATE OF ANALYSIS

STP pHox Plus/L/C/M Sodium Sensor
Lot Number 25289013

Ref Number 27810
Exp Date 2027-04-16
Date of Manufacture 2025-10-16
UDI# (01)00385480278102(11)251016(17)270416(10)25289013

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.00 - 10.50	Pass
Std C mV	-80.00 - +80.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector

Title

11/5/2025

Date