



CERTIFICATE OF ANALYSIS

STP pHox Ultra™/ CCX BUN Sensor
Lot Number 25296015

Part Number 48555
Exp Date 2027-01-23
Date of Manufacture 2025-10-23
UDI# (01)00385480485555(11)251023(17)270123(10)25296015

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	≥ 9.60	Pass
pHOx Ctl L4	14 - 24	Pass
pHOx Ctl L5	49 - 63	Pass
Std D mV	0.00 ± 60.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

11/13/2025

Date