



**CERTIFICATE OF COMPLIANCE Page 1/2**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.	
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE, PACKAGED	
Catalog Number:	47940	
Nova Biomedical P/N	47940	
Target Range (mg/dL)	SD (mg/dL) /CV (%)	
30-50	≤ 4 mg/dL	PASS / FAIL
51-110	≤ 6.0%	PASS / FAIL
111-150	≤ 5.0%	PASS / FAIL
151-250	≤ 5.0%	PASS / FAIL
251-400	≤ 5.0%	PASS / FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/-12% of reference method ≥ 95%		PASS / FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/-15% of reference method ≥ 98%		PASS / FAIL
Qty:	304	
Lot No.	2125195309	
Mfg. Date:	2025.07.14	
Expiration Date:	2028.01.14	

Maureen Brown  
 Quality Control Manager

07/25/25  
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454



**ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2**  
**RELEASE TEST RESULTS**

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE
Catalog Number:	47940
Nova Biomedical P/N	47940
Lot No.	2125195309

(Spiked whole blood) Target Range:30-50		
Mean (mg/dL) N=100	SD (mg/dL) ≤4	YSI (reference, mg/dL)
41	2.0	43.4
(Spiked whole blood) Target Range:51-110		
Mean (mg/dL) N=100	CV% ≤6.0	YSI (reference, mg/dL)
102	3.0	103.3
(Spiked whole blood) Target Range:111-150		
Mean (mg/dL) N=100	CV% ≤5.0	YSI (reference, mg/dL)
140	3.2	140.0
(Spiked whole blood) Target Range:151-250		
Mean (mg/dL) N=100	CV% ≤5.0	YSI (reference, mg/dL)
222	3.4	220.0
(Spiked whole blood) Target Range:251-400		
Mean (mg/dL) N=100	CV% ≤5.0	YSI (reference, mg/dL)
357	3.1	351.0

Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method ≥ 95%	Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98%
100.0%	100.0%

*Marc Broome*  
 Quality Control Manager Date 7/25/25  
 Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454