

***nova***  
*biomedical*  
**CERTIFICATE OF COMPLIANCE Page 1/2**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.	
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE, PACKAGED	
Catalog Number:	47940	
Nova Biomedical P/N	47940	
Target Range (mg/dL)	SD (mg/dL) /CV (%)	
30-50	≤ 4 mg/dL	<u>PASS</u> / FAIL
51-110	≤ 6.0%	<u>PASS</u> / FAIL
111-150	≤ 5.0%	<u>PASS</u> / FAIL
151-250	≤ 5.0%	<u>PASS</u> / FAIL
251-400	≤ 5.0%	<u>PASS</u> / FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/-12% of reference method ≥ 95%		<u>PASS</u> / FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/-15% of reference method ≥ 98%		<u>PASS</u> / FAIL
Qty:	306	
Lot No.	2125300309	
Mfg. Date:	2025.10.27	
Expiration Date:	2028.04.27	

Mark Broome  
 Quality Control Manager

11/6/25  
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454

Released Confidential



**ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2**  
**RELEASE TEST RESULTS**

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE
Catalog Number:	47940
Nova Biomedical P/N	47940
Lot No.	2125300309

(Spiked whole blood) Target Range:30-50		
Mean (mg/dL) N=100	SD (mg/dL) ≤4	YSI (reference, mg/dL)
50	3.3	48.9
(Spiked whole blood) Target Range:51-110		
Mean (mg/dL) N=100	CV% ≤6.0	YSI (reference, mg/dL)
106	3.4	99.7
(Spiked whole blood) Target Range:111-150		
Mean (mg/dL) N=100	CV% ≤5.0	YSI (reference, mg/dL)
145	3.8	138.0
(Spiked whole blood) Target Range:151-250		
Mean (mg/dL) N=100	CV% ≤5.0	YSI (reference, mg/dL)
241	3.8	226.3
(Spiked whole blood) Target Range:251-400		
Mean (mg/dL) N=100	CV% ≤5.0	YSI (reference, mg/dL)
371	3.6	351.5

Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method ≥ 95%	Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98%
96.6%	98.8%

Marie Broome 11/6/25  
 Quality Control Manager Date  
 Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454