



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, STATSENSOR™ CREATININE, 50 COUNT
Catalog Number:	43370
Nova Biomedical P/N	43370
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS/FAIL PASS/FAIL PASS/FAIL
Qty:	4,986
Lot No.	4925258129
Mfg. Date:	2025-09-15
Expiration Date:	2026-09-15
Unique Device Identifier (UDI):	(01)20385480433703 (11)250915 (17)260915 (10)4925258129

Marc Boome
 Quality Control Manager

09/24/25
 Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454