



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	TEST STRIPS, STATSENSOR™ CREATININE, 50 COUNT
Catalog Number:	43370
Nova Biomedical P/N	43370
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS/FAIL PASS/FAIL PASS/FAIL
Qty:	3,816
Lot No.	4925259129
Mfg. Date:	2025-09-16
Expiration Date:	2026-09-16
Unique Device Identifier (UDI):	(01)20385480433703 (11)250916 (17)260916 (10)4925259129

M. Broome
 Quality Control Manager

9/24/2025
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454