




**CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	M/D
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	<del>PASS</del> / FAIL <del>PASS</del> / FAIL <del>PASS</del> / FAIL
Qty:	3,841
Lot No.	4925260129
Mfg. Date:	2025-09-17
Expiration Date:	2026-09-17
Unique Device Identifier (UDI):	(01)20385480432720 (11) <u>2 5 0 9 1 7</u> (17) <u>2 6 0 9 1 7</u> (10) <u>4 9 2 5 2 6 0 1 2 9</u>

  
 \_\_\_\_\_  
 Quality Control Manager Date 9/26/25

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454