



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	TEST STRIPS, NOVA MAX PRO CREATININE eGFR, 50 COUNT
Catalog Number:	63983
Nova Biomedical P/N	63983
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10%	PASS/FAIL PASS/FAIL
Qty:	612
Lot No.	2730225217
Mfg. Date:	2025-08-05
Expiration Date:	2026-08-05
Unique Device Identifier (UDI):	(01)20385480639839 (11)250805 (17)260805 (10)2730225217

Marc Broome
 Quality Control Manager

8/18/2025
 Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454