



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Chemistry Controls
Auto-Cartridge 6/Day
Lot Number 25302061

Ref Number 48905
Exp Date 2027-04-21

Date of Manufacture 2025-10-21

UDI# (01)00385480489058(11)251021(17)270421(10)25302061

Table with 6 columns: Analyte, Level 4 Spec, Level 4 Pass/Fail, Level 5 Spec, Level 5 Pass/Fail. Rows include Na+, K+, Cl-, Ca++, Mg++, Glu, Lac, and BUN.

Table for Level 6 QC Performance Check. Columns: Analyte, Spec, Pass/Fail. Row: Glu mg/dL, 700 - 1500, Pass.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Handwritten signature

Approval

QC Inspector
Title

11/13/2025
Date