nova biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	TEST STRIPS, NOVA VET KETONE, 25 CT., PACKAGED
Catalog Number:	52587
Nova Biomedical P/N	52587
Target Range / MPE / SD/CV (%) 0.1-0.8 mmol/L / +- 0.3 mmol/L / ≤ 0.1 mmol/L 1.5-2.5 mmol/L / +- 20% / ≤ 10% 3.5-5.0 mmol/L / +- 20% / ≤ 10%	(PASS / FAIL) (PASS / FAIL) (PASS / FAIL)
Qty:	1,152
Lot No.	1760425231
Mfg. Date:	2025-08-19
Expiration Date:	2027-08-19
Unique Device Identifier (UDI):	(01)20385480525873 (11)250819 (17)270819 (10)1760425231

Sr Mfg. Manager — Quality Control Manager

9/12/25 — Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454

Released Confidential

Printed on: 02 Sep 2025, 07:15:06 am; Printed by: CCHAP100746.