



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1

Nova Reference 54520
LifeScan Reference 02315402
Lot Number 0425307241

Date of Manufacture 2025-11-03

Exp Date 2027-11-03

UDI# (01)00385480545204(11)251103(17)271103(10)0425307241

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	61

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink that reads "Oeise Rammer".

Approval

QC Inspector

Title

11/19/2025

Date