



**CERTIFICATE OF ANALYSIS**

Stat Profile pHox Ultra / CCX SO2 Sensor  
Lot Number 25321019

Part Number 48591  
Exp Date 2028-11-17  
Date of manufacture 2025-11-17  
UDI#(01)00385480485913(11)251117(17)281117(10)25321019

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	7.00 - 14.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector  
Title

11/19/2025  
Date