



**CERTIFICATE OF ANALYSIS**

Prime Reference Cartridge  
Lot Number 25310030

Part Number 42043

Exp Date 2027-05-06

Date of Manufacture 2025-11-06

UDI# (01)00385480420433(11)251106(17)270506(10)25310030

	<u>Spec</u>	<u>Pass/Fail</u>
Na+ Slope	9.50 - 12.00	Pass
K+ Slope	9.00 - 11.50	Pass
Hct Slope	2.00 - 3.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Lumana Ahtu".

Approval

QC Inspector

Title

11/20/2025

Date