



## CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 pH/Gas Reference Sensor  
Lot Number 25318017

Part Number 57467  
Exp Date 2027-05-14  
Date of Manufacture 2025-11-14

|                | <u>Spec*</u>  | <u>Pass/Fail</u> |
|----------------|---------------|------------------|
| pH Slope       | 8.70 - 11.80  | Pass             |
| BioPro Ctl L 1 | 6.628 - 6.728 | Pass             |
| BioPro Ctl L 3 | 7.306 - 7.406 | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

\* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in black ink, appearing to be "John", is written over a horizontal line.

Approval

QC Inspector  
Title

12/01/2025  
Date