



CERTIFICATE OF ANALYSIS

STP Prime Plus VET Cal-Cartridge 500 Sample
Lot Number 25322081

Ref Number 58396
Exp Date 2027-05-03
Date of Manufacture 2025-11-03

Table with 3 columns: Cal A L/N 25308037, Cal B L/N 25308038, Cal C L/N 25308039. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN mmol/L and tHb g/dL with Spec, Pass/Fail, and Result columns.

Table with 3 columns: Cal D L/N 25308040, Flush L/N 25307027, Reference L/N 25308010. Rows include pH @ 37C, PO2, KCl with Spec, Pass/Fail, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Handwritten signature: Olivia Rammer

Approval

QC Inspector
Title

12/2/2025
Date