



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 100 Sample
Lot Number 25322081

Ref Number 57825
Exp Date 2027-05-03
Date of Manufacture 2025-11-03
UDI# (01)00385480578257(11)251103(17)270503(10)25322081

Table with 3 columns: Cal A L/N 25308037, Cal B L/N 25308038, Cal C L/N 25308039. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, and tHb g/dL with Spec, Pass/Fail, and Result columns.

Table with 3 columns: Cal D L/N 25308040, Flush L/N 25307027, Reference L/N 25308010. Rows include pH @ 37C, PO2, and KCl with Spec, Pass/Fail, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Handwritten signature: Olivia Rammer

Approval

QC Inspector
Title

12/2/2025
Date