



CERTIFICATE OF ANALYSIS

Bioprofile Control Level 5
Lot Number 530826

Catalog Number 24619
Exp Date 2027-05

	<u>Spec</u>	<u>Pass/Fail</u>
Gln mmol/L	4.60 - 5.40	Pass
Glu mmol/L	1.80 - 2.20	Pass
Gluc g/L	9.60 - 10.60	Pass
Lac g/L	4.60 - 5.40	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP) and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Lac, Gluc.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "M. Houghton", with a long, sweeping flourish extending to the right.

Approval

QC Inspector

Title

12/1/2025

Date