



CERTIFICATE OF ANALYSIS

Prime Reference Cartridge
Lot Number 25322037

Part Number 42043
Exp Date 2027-05-18
Date of Manufacture 2025-11-18
UDI# (01)00385480420433(11)251118(17)270518(10)25322037

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-----------|--------------|------------------|
| Na+ Slope | 9.50 - 12.00 | Pass |
| K+ Slope | 9.00 - 11.50 | Pass |
| Hct Slope | 2.00 - 3.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Z. V. [Signature]

Approval

QC Inspector

Title

12/2/2025

Date