



CERTIFICATE OF ANALYSIS

Bioprofile Flex Linearity Solution HIGH
Lot Number 25303041

Catalog Number 45279
Exp Date 2027-04

	<u>Spec</u>	<u>Pass/Fail</u>
Gln mmol/L	4.60 - 5.40	Pass
Glu mmol/L	4.60 - 5.40	Pass
Gluc g/L	13.50 - 16.50	Pass
Lac g/L	4.60 - 5.40	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Gluc, Lac

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "M. Houghton", is written over a horizontal line.

Approval

QC Inspector

Title

12/8/2025

Date