



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 25317007

Part Number 48554
Exp Date 2027-02-13
Date of Manufacture 2025-11-13
UDI# (01)00385480485548(11)251113(17)270213(10)25317007

	<u>Spec*</u>	<u>Mean Result</u>
Slope	≥ 15.00	37.10
pHox Ultra/CCX Ctl L4	1.7 - 2.3	1.8
pHox Ultra/CCX Ctl L5	6.3 - 7.7	7.0
Lin Soln Set A Level 1	14.0 - 20.0	18.7

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Z. V. Admwal

Approval

QC Inspector

Title

11/21/2025

Date