



**CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS / FAIL PASS / FAIL PASS / FAIL
Qty:	2,417
Lot No.	4925317129
Mfg. Date:	2025-11-13
Expiration Date:	2026-11-13
Unique Device Identifier (UDI):	(01)20385480432720 (11) 251113 (17) 261113 (10) 4925317129

Mar Broome  
Quality Control Manager

12/2/25  
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454