



CERTIFICATE OF ANALYSIS

Allegro HbA1c Test Cartridge
Lot Number 25300001

Ref Number 54215
Exp Date 2026-10-27
Date of Manufacture 2025-10-27
UDI# (01)00385480542159(11)251027(17)261027(10)25300001

Ctl Level 1			Ctl Level 2		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HbA1c %	4.6 - 6.8	Pass	HbA1c %	6.7 - 11.1	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

11/20/2025
Date