



CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Ultra/CCX pO2 Sensor
Lot Number 25321025

Ref Number 48598

Exp Date 2027-11-17

Date of Manufacture 2025-11-17

UDI# (01)00385480485982(11)251117(17)271117(10)25321025

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	-3.00 to -6.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "John", written in a cursive style.

Approval

QC Inspector
Title

12/3/2025
Date