



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 200 Sample
Lot Number 25330031

Ref Number 53360
Exp Date 2027-05-11
Date of Manufacture 2025-11-11
UDI# (01)00385480533607(11)251111(17)270511(10)25330031

Cal A	L/N	25315061	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.353	N/A
pCO2 mmHg		25.0	N/A

Cal B	L/N	25315062	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	25316028	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		51.9	N/A
pO2 mmHg		103.3	N/A

Reference	L/N	25316012	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/15/2025
Date