



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 300 Sample
Lot Number 25330031

Ref Number 53359
Exp Date 2027-05-11
Date of Manufacture 2025-11-11
UDI# (01)00385480533591(11)251111(17)270511(10)25330031

Cal A	L/N	25315061		
		<u>Assigned Value*</u>	<u>Result</u>	
	pH	7.353	N/A	
	pCO2 mmHg	25.0	N/A	

Cal B	L/N	25315062		
		<u>Spec</u>	<u>Pass/Fail</u>	
	pH	6.837 - 6.843	Pass	
	pO2 mmHg	167.0	N/A	

Flush	L/N	25316028		
		<u>Assigned Value*</u>	<u>Result</u>	
	pCO2 mmHg	51.9	N/A	
	pO2 mmHg	103.3	N/A	

Reference	L/N	25316012		
		<u>Spec</u>	<u>Pass/Fail</u>	
	KCl	1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/15/2025
Date