CERTIFICATE OF ANALYSIS

Stat Profile pHOx/Basic Calibrator Cartridge A
Lot Number 25325080

Ref Number 23935
Exp Date 2027-05-13
Date of Manufacture 2025-11-13
UDI# (01)00385480239356(11)251113(17)270513(10)25325080

Cal A L/N 25321035

	Spec	Pass/Fail
HCO3 mmol/L	20.1 - 20.9	Pass
Na mmol/L	139.3 - 141.3	Pass

Cal B L/N 25321036

	Spec	Pass/Fail
HCO3 mmol/L	40.2 - 41.8	Pass

Cal C L/N 25321037

	Spec	Pass/Fail
pH @ 37C	7.342 - 7.348	Pass

Cal D L/N 25321038

		Spec	Pass/Fail
pH	@ 37C	6.792 - 6.798	Pass
Na	mmol/L	71.5 - 72.5	Pass

Reference L/N 25317032

	Spec	Pass/Fail
KCl mol/L	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi.M

Approval

QC Inspector
Title

12/15/2025
Date