



CERTIFICATE OF ANALYSIS

StatStrip Glucose Control Solution Level 3

Nova Ref Number 54522
LifeScan Ref Number 02315602
Lot Number 0425324243

Date of Manufacture 2025-11-20
Exp Date 2027-11-20

UDI # (01)00385480545228(11)251120(17)271120(10)0425324243

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	304

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/12/2025
Date