



CERTIFICATE OF ANALYSIS

Stat Profile pHox Ultra/CCX Calibrator Cartridge 1
Lot Number 25329040

Ref Number 48834
Exp Date 2027-05-14

Date of Manufacture 2025-11-14

UDI# (01)00385480488341(11)251114(17)270514(10)25329040

Table with 3 columns: Cal-A L/N 25321034, Cal-B L/N 25318030, Cal-C L/N 25318031, Cal-D L/N 25318032, Flush L/N 25324012, Reference L/N 25318010. Each column contains test results for various analytes (pH, Na, K, Cl, Ca, Mg, Glu, Lac, BUN, HCO3) with Spec, Pass/Fail, and values.

This certifies that this product was manufactured and tested at Nova BiomedCal-Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 MediCal-Devices Quality Management Systems Requirements, MediCal-Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Handwritten signature: Auishi, M

Approval

QC Inspector
Title

12/11/2025
Date