



CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX Ionized Mg Sensor
Lot Number 25330022

Part Number 48556
Exp Date 2027-02-26
Date of Manufacture 2025-11-26
UDI# (01)10385480485569(11)251126(17)270226(10)25330022

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.70 - 14.00	Pass
Std C mV	-100.00 - +60.00	Pass
pHOx Ctl L4	0.55 - 0.73	Pass
pHOx Ctl L5	1.00 - 1.30	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Z. Veramwala

Approval

QC Inspector

Title

12/15/2025

Date