



**CERTIFICATE OF ANALYSIS**

STP pHox Plus/L/C Chloride Sensor  
Lot Number 25316022

Part Number 27812  
Exp Date 2027-02-12  
Date of Manufacture 2025-11-12

UDI# (01)10385480278123(11)251112(17)270212(10)25316022

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	≥ 9.00	Pass
Std A mV	55.00 - 95.00	Pass
pHOx Ctl L1	119.9 - 128.9	Pass
pHOx Ctl L3	78.2 - 88.2	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

12/16/2025

Date