



CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX Lactate Sensor
Lot Number 25336020

Part Number 48553
Exp Date 2027-12-02
Date of Manufacture 2025-12-02
UDI# (01)00385480485531(11)251202(17)271202(10)25336020

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	77 - 91	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "M. Knight", with a long, sweeping flourish extending to the right.

Approval

QC Inspector
Title

12/17/2025
Date