



CERTIFICATE OF ANALYSIS

Prime Reference Cartridge
Lot Number 25297035

Part Number 42043
Exp Date 2027-04-24
Date of Manufacture 2025-10-24
UDI# (01)00385480420433(11)251024(17)270424(10)25297035

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-----------|--------------|------------------|
| Na+ Slope | 9.50 - 12.00 | Pass |
| K+ Slope | 9.00 - 11.50 | Pass |
| Hct Slope | 2.00 - 3.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "G. F. ...".

Approval

QC Inspector
Title

12/15/2025
Date