



CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX Calcium Sensor
Lot Number 25316020

Part Number 48581
Exp Date 2027-02-12
Date of Manufacture 2025-11-12
UDI# (01)00385480485814(11)251112(17)270212(10)25316020

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	8.50 - 12.50	Pass
Std C mV	0.00 ± 50.00	Pass
pHOx Ctl L4	1.03 - 1.19	Pass
pHOx Ctl L5	1.51 - 1.75	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

12/4/2025

Date