



**CERTIFICATE OF ANALYSIS**

STAT PROFILE pHox Ultra/CCX Potassium Sensor  
Lot Number 25330025

Ref Number 48582

Exp Date 2027-02-26

Date of Manufacture 2025-11-26

UDI# (01)00385480485821(11)251126(17)270226(10)25330025

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.00 to 11.00	Pass
Std C mV	0 ± 50.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "Lumana Ahtu".

Approval

QC Inspector  
Title

12/22/2025  
Date