



CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pO2 Sensor
Lot Number 25336036

Ref Number 21521
Exp Date 2027-12-02

Date of Manufacture 2025-12-02

UDI# (01)00385480215213(11)251202(17)271202(10)25336036

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	-3.00 to -6.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in cursive script, appearing to read "Lumina Ahtu".

Approval

QC Inspector
Title

12/22/2025
Date