



**CERTIFICATE OF ANALYSIS**

STP pHox Ultra / CCX pH Sensor  
Lot Number 25336034

Part Number 48580  
Exp Date 2027-06-02  
Date of Manufacture 2025-12-02  
UDI# (01)00385480485807(11)251202(17)270602(10)25336034

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.20 - 11.20	Pass
Std C mV	0.00 ± 110.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Z. Verdamwal*

Approval

QC Inspector

Title

12/22/2025

Date