CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 2

Nova Reference	54521	Date of Manufacture	2025-11-18
LifeScan Reference	02315502	Exp Date	2027-11-18
Lot Number	0425322242	UDI#	(01)00385480545211(11)251118(17)271118(10)0425322242

	Spec	Mean Result
Glucose mg/dL	105 - 125	112

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/23/2025
Date