CERTIFICATE OF ANALYSIS

Stat Profile pHOx Ultra / CCX SO2 Sensor
Lot Number 25350033

Part Number 48591
Exp Date 2028-12-16
Date of manufacture 2025-12-16
UDI#(01)00385480485913(11)251216(17)281216(10)25350033

	Spec	Pass/Fail
Slope	7.00 - 14.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/23/2025
Date