



**CERTIFICATE OF ANALYSIS**

STP pHox Plus/L/C/M Potassium Sensor  
Lot Number 25316027

Part Number 27811  
Exp Date 2027-02-12

Date of Manufacture 2025-11-12

UDI# (01)00385480278119(11)251112(17)270212(10)25316027

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	≥ 9.60	Pass
Std A mV	0 ± 60.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

12/23/2025

Date