



CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Ultra/CCX pCO2 Sensor
Lot Number 25346019

Ref Number 48600
Exp Date 2027-12-12
Date of Manufacture 2025-12-12
UDI# (01)00385480486002(11)251212(17)271212(10)25346019

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.40 - 11.60	Pass
Std C mV	0.00 ± 150.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/31/2025
Date