CERTIFICATE OF ANALYSIS

STP Prime+ Vet Cal-Cartridge 500 Sample w. Creat/BUN
Lot Number 25349041

Ref Number 58404
Exp Date 2027-04-15
Date of Manufacture 2025-10-15

Cal A L/N 25289016

	Spec	Pass/Fail
HCO3 mmol/L	10.78 - 11.22	Pass
Na mmol/L	134.6 - 136.4	Pass
K mmol/L	3.97 - 4.03	Pass
Cl mmol/L	107.1 - 108.9	Pass
Ca mmol/L	1.02 - 1.06	Pass
Mg mmol/L	0.49 - 0.51	Pass
Glu mg/dL	78.8 - 81.2	Pass
Lac mmol/L	1.8 - 2.2	Pass
BUN mg/dL	9.75 - 10.25	Pass
Creat mg/dL	0.98 - 1.02	Pass

Cal B L/N 25289017

	Spec	Pass/Fail
pH	6.831 - 6.837	Pass
Na mmol/L	76.5 - 77.5	Pass
K mmol/L	9.92 - 10.08	Pass
Ca mmol/L	1.89 - 1.95	Pass
Mg mmol/L	1.48 - 1.52	Pass

	*Assigned Value**	*Result*
tHb g/dL	*15.6*	*N/A*

Cal C L/N 25289018

	Spec	Pass/Fail
HCO3 mmol/L	21.56 - 22.44	Pass
Cl mmol/L	68.4 - 69.6	Pass
Glu mg/dL	197 - 203	Pass
Lac mmol/L	9.0 - 11.0	Pass
BUN mg/dL	49.4 - 50.6	Pass
Creat mg/dL	4.9 - 5.1	Pass

Cal D L/N 25289019

	Spec	Pass/Fail
pH @ 37C	7.351 - 7.357	Pass

Flush L/N 25303036

	*Assigned Value**	*Result*
PO2	*103.3*	*N/A*

Reference L/N 25289020

	Spec	Pass/Fail
KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Anishi.M

Approval

QC Inspector
Title

12/23/2025
Date