



**CERTIFICATE OF ANALYSIS**

Allegro UACR Test Cartridge  
Lot Number 25344005

Ref Number 54214

Exp Date 2026-12-10

Date of Manufacture 2025-12-10

UDI# (01)00385480542142(11)251210(17)261210(10)25344005

	Ctl Level 1			Ctl Level 2	
	Spec	Pass/Fail		Spec	Pass/Fail
Albumin mg/L	36-60	Pass	Albumin mg/L	135-211	Pass
Creatinine mg/dL	61-105	Pass	Creatinine mg/dL	175-291	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

12/23/2025  
Date