



CERTIFICATE OF ANALYSIS

Stat Profile pHox Sodium Sensor
Lot Number 25357055

Ref Number 21523
Exp Date 2026-12-23
Date of Manufacture 2025-12-23
UDI# (01)00385480215237(11)251223(17)261223(10)25357055

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.00 - 11.00	Pass
Std A mV	10.00 - 130.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink that reads "Julie Ramona".

Approval

QC Inspector
Title

1/12/2026
Date